



BG 2100

ISE Electrolyte & Blood Gas Analyzer

- Measured Parameters: PCO₂, PO₂, pH, K⁺, Na⁺, Cl⁻, Ca⁺⁺
- 240 x 128 dots LCD Display
- Soft Keypad
- ≤200 µL Sample Volume – Whole Blood, Serum, Plasma
- 65 Samples / hour Throughput
- Built-in Thermal Printer



BG 2100 is an innovative ISE Electrolyte and Blood Gas Analyzer complementing our existing ISE NG. The analyzer measures K⁺, Na⁺, Cl⁻, Ca⁺⁺, pH, PCO₂ and PO₂ in whole blood, serum and plasma. The analyzer has a compact design, is easy to operate, economical and is virtually maintenance free. The aspirate probe has a self-cleansing mechanism affording enhanced operator safety, low carry-over and cross-contamination.

Gas Specifications

- Low Gas Composition
CO₂ 5% O₂ 20% N₂ 75%
- High Gas Composition
CO₂ 10% N₂ 90%
- Tank(L) Pressure: 15MPa
- Tank(S) Pressure: 10Mpa
- Two Level Gas Regulator

Test Ranges

Channel	Linearity	CV	SD
pH(cH)	6.800–8.000 (158.5 –10 mmol/L)	≤0.1%	20 * 10 ⁻³
PCO ₂	1.607–26.67 kPa (8.0–200.0 mmHg)	≤1.5%	≤4% or 0.13kPa
PO ₂	0–106.7 kPa (0–800mmHg)	≤1.0%	≤3% or 0.13kPa
K ⁺	2.0–10 mmol/L	≤2.0%	≤0.4 mmol/L
Na ⁺	50–200 mmol/L	≤1.0%	≤4 mmol/L
Cl ⁻	50–200 mmol/L	≤2.0%	≤4 mmol/L
Ca ⁺⁺	0.5–3.0 mmol/L	≤3.0%	≤0.3 mmol/L

Technical Specifications

- Equipment Interface: 240 × 128 dots LCD display with soft key pad
- Analysis principle: Direct "Flow Thru" electrodes
- Sample volume: ≤200 µl
- Sample type: Whole Blood, Serum, Plasma,
- Sample Application: Sample Cup, Capillary or Collection Tube
- Throughput: 65 samples/hour
- Calibration: Automatic two-point calibration with additional option for carrying out one-point and two-point calibration manually
- Pre adjusted Standby mode with automatic calibration steps
- Quality Control: 2 different levels of quality controls
- Printer: Built in Stationary Head Graphical Thermal Printer
- Storage Capacity: 1000 samples
- Interface port: RS 232
- Power: 110 ~ 230 V AC, 50/60 Hz
- Dimensions: 452(H) × 452(W) × 500(D) mm approx.
- Equipment weight: 15 Kg approx.
- Operating environment: 15° to 30°C, 20% ~ 80% RH, 86 ~ 106 kPa

Parameter

Model BG 2100

Measured Parameters

PCO₂, PO₂, pH, K⁺, Na⁺, Cl⁻, Ca⁺⁺

Calculated Parameters

pH(TC), PCO₂(TC), PO₂(TC), HCO₃⁻, SBC, AG, BE, BEecf, TCO₂, sO₂%, A-aDO₂, RI, Ca⁺⁺(7.4), TCa⁺⁺



Analytical[®]
Technologies Limited

Analytical Group Companies
Analytical Technologies Limited
Analytical Bio-Technologies
Analytical Bio- Med
HPLC Solutions

Corporate & Regd. Office :
Analytical House, # E67 & E68, Ravi Park, Opp., Sainath Park, Vasna Road,
Baroda, Gujarat - 390015 INDIA. Tele : 0265-2253620 / 2252839
Fax : 0265-2254395
E-mail : info@hplctechnologies.com/info@ais-india.com
Web : www.ais-india.com / www.analyticalgroup.net