



# GC-MS-3068 Gas Chromatograph Mass Spectrometer 3068

- High performance, high reliabillity
- Low cost for customers of all types
- Holding multiple patents



EPC / PRODUCTS / APPLICATION / SOFTWARE / ACCESSORIES / CONSUMABLES / SERVICES

**Analytical Technologies Limited** 

An ISO 9001 Certified Company www.analyticalgroup.net





Gas Chromatograph - Mass Spectrometer 3068 GCMS 3068 -- High performance-to-price ratio





#### Introduction:

ATL'S new generation Gas Chromatograph mass spactrometer (GC-MS) system offers high performance, high reliability, and low cost for customers of all types. Holding multiple patents ATL manufactures high quality GC-MS 3068 which is widely used in industrical inspection, food safety, environmental protection, etc.

#### Hardware:

Electronic pressureflow control system (EPC/EFC) on GC 3068improves baseline stability and sensitivity. Patented EI filaments provide high efficiency of electron emission.

Pre-quad removes undesirable contamintion before the main analytical quadrupote and decreases cleaning frequency

High quility vacuum system combined with high energy dynode (HED) electron multiplier (EM) offer magnificent sensitivity

Self-protevtor safeguards the system against the situations, which would normally damage expensive components and require significant repair.

Advanced digital compensation technology on Radio Frequency (RF) power supply guarantees satisfactory sensivity and resolution over full mass range.

#### Analchrom is the workstation software for GC-MS 3068 it features

a user-friendly interface to simultaneously control autosampler, chromatograph and mass spectrometer. Using high speed net card. Software have Security, Audit trail, System check, Software integrity and system Suitability test should as standard functions. Flexible report Format for Method, chromatogram, Mass Spectrum, Peak table, Quantitation result, calibration curve, Status Log, texts, graphics. Automated tuning & File management functions with Library Search facility.

In INDIA, ATL has been the only authorized distributor of NIST database (System have retention time locking facility, molecular Structure etc.) which is the most popular for mass spectral search and examination of unknowns and target. Compounds in qualitative and quantitative analysis Complete Software control of vacuum system with Auto Start-up / Shut-down and vacuum protection against Power Failures

#### **Turbo Molecular Pump**

Turbo Molecular pump with capacity of 250L/Sec shall be provided









# **Specification:**

| Temperature program steps            | 24 steps   |  |  |
|--------------------------------------|--|--|--|
| Heating zone                         | 6  |  |  |
| Display                              | Touch Screen   |  |  |
| Capability to control electronically | 16 channels for gas  |  |  |
| Capability to install                | 5 detectors +1 MS detector + 3 sample injectors with independent temperature control |  |  |
| Channel analysis                     | 3  |  |  |
| Data acquisition speed               | ≤ 3ms (250Hz) for all modules  |  |  |
| Pressure and detector gas            | digitally controlled   |  |  |
| Memory                               | able to store 14 methods   |  |  |
| Safety feature                       | Memory Protection When Power Off, Leakage Detection, Power line Failure              |  |  |
| Self diagnostic                      | provided   |  |  |
| Pressure                             | 0 to 150 PSI   |  |  |
| Operating Temperature range of inlet | 50°C to 450°C in 1°C increments  |  |  |
| Injection volume                     | Up to 150 μL   |  |  |
| Retention time repeatability         | <0.0008 min  |  |  |
| Peak area repeatability              | <1 % RSD   |  |  |

# **Injection Port:**

Independently temperature controlled injector units are provided Injection port unit: Split/split less injection unit provided as standard injection ports shall be heated simultaneously

# **Oven Specfications**



| Temperature range Ambient +4°C ~ 4             | 450°C  |  |  |  |
|--|--|--|--|--|
| Column oven dimension                          | 278x310x165mm=15L; accommodate up to<br>2 column 105m x 0.53mm ID capillary column |  |  |  |
| Oven volume:                                   | 15 Liters  |  |  |  |
| Temperature calibration at                     | 0.04°C   |  |  |  |
| Temperature variation coefficient:             | 0.040°C/0°C  |  |  |  |
| Programming rate setting:                      | 130°C/min  |  |  |  |
| Cooling rate:                                  | 450°C to 50°C within 1 minute  |  |  |  |
| Temp. accuracy                                 | 0.1°C  |  |  |  |
| Oven temperature programmed rate setting range | -250 to 250°C.   |  |  |  |
| Total system time:                             | ~9999.99 min   |  |  |  |
| Temperature set point resolution:              | ≤0.1°C   |  |  |  |
| Temperature deviation:                         | ≥ 2℃   |  |  |  |
| Ramp rate                                      | 120°C/min  |  |  |  |
| Heating ramp                                   | 20 ramps   |  |  |  |
| Negative ramps are allowed                     |  |  |  |  |

# **Mass Spectrometry Specifications**

| wass spectrometry specifications        |  |  |  |  |
|---|--|--|--|--|
| Mass range:                             | M/z 1 to 1300  |  |  |  |
| Mass Stability                          | ±0.1 amu / 48 hours  |  |  |  |
| Resolution:                             | R=1M (FWHM)  |  |  |  |
| Scan rate:                              | 20000 amu/sec max.   |  |  |  |
| Direct connection with capillary column |  |  |  |  |
| Temperature:                            | room temperature to 450°C  |  |  |  |
| lon source:                             | up to 450°C  |  |  |  |
| Dynamic range :                         | 10e6   |  |  |  |
| Maximum flow of He to MS :              | 18 ml/min  |  |  |  |
| Stabilizer:                             | 0.1 amu/48 hour  |  |  |  |
| Temperature:                            | 50 ~ 450°C   |  |  |  |
| Filament                                | Dual   |  |  |  |
| Sensitivity (Signal/Noise)              | El scan: 3000:1 S/N for 1 pg OFN   |  |  |  |
| MDL                                     | 8 sequential 10 fg OFN Split less injections<br>monitored, chromatographic peak area with 99%<br>confidence interval: IDL ≤ 5 fg |  |  |  |
| Electron Ionization Voltage             | 10-200 eV  |  |  |  |
| Electron Ionization Current             | 5-250 μΑ   |  |  |  |
| Detector                                | Electron Multiplier  |  |  |  |
| Ionization Mode                         | Electron Ionization  |  |  |  |
| Mass Analyzer                           | Single Quadruple   |  |  |  |



#### **Detectors:**


| Detector | Max<br>operating<br>temp. | Limit of detection           | Baseline<br>noise        | Baseline drift<br>(after 2hrs<br>stabilization) | Linear<br>dynamic<br>range | Data acquisition rate |
|----------|---------------------------|------------------------------|--------------------------|---|----------------------------|-----------------------|
| FID      | 450°C                     | ≤1.4 pgC/s                   | <2 x 10 <sup>-14</sup> A | 5 x 10 <sup>-14</sup> A/30min                   | ≥10 <sup>7</sup>           | > 300Hz               |
| ECD      | 450°C                     | ≤3 x 10 <sup>-14</sup> pg/ml | <20uV                    | < 50uV /30min                                   | ≥10 <sup>4</sup>           |                       |

Flameout detection and re-ignition shall be possible

# **Example:**

#### **Application:**

Industrial Inspection
Food safety and hygiene
Environmental Protection
Pharmaceutical development
Criminal investigation/Forensics



Poly Brominated Biphenyis (PBBs)/Poly Brominated Diphenyl Ethers (PBDEs)

Mass spectrum of deca polytrominated biphemyl ethers

# Auto sampler:

- Vial capacity: 2 mL (Optional micro-volume vials: 300 μL)
- Big displayer with double -tower automatic injection;
- Tray vial quantity: 16 or 150;
- Injection volume: 0.1-100uL;
- Sampling accuracy:+0.01uL;
- Injection Port: split/splitless capillary
- Sample injection method: Liquid sample injection
- Injection needle: 5, 10, 50, 100uL;
- Injection loop: multiport(0.25ml, 0.5ml and 1ml)
- Injection repeatability: <0.5%</li>
- Maximum Temperature: 450°C
- Cross contamination: Less than 10<sup>-4</sup>
- Temperature control range: RT+5°C~ 450°C (0.1°C)
- Maximum Pressure: 0-150 psi (with EPC)
- Vial volume: 10 ml to 22 ml
- Transfer line based with loop system for precise quantification
- Transfer Line to the GC temperature range: 50 to 200 Degree
- Automatic leak check and gas saving facility
- System heating up to 200°C or better in increments of 1°C with shaker
- 100 sample vial capacity

#### Libraries:

Latest NIST Library, Wiley, Pesticide Library, FFNSC Library etc.

# Split/Split less:

Packed purge injection port (PPIP)

Split/splitless capillary port (S/SL)

- 1) Max Temperature: 450"
- 2) Split ratio: 12500:1
- 3) Gas saver mode to reduce gas consumption without compromising performance
- 4) Electronic septum purge flow control to eliminate carry-over
- 5) User-installable within a few minutes.

# >>> Regulatory compliances



#### **>>>** Corporate Social Responsibility



Analytical Foundation is a Nonprofit Organization (NGO) found for the purpose of:

- 1.Research & Innovation Scientist's awards / QC Professional Award: Quality life is possible by innovation only and the innovation is possible by research only, hence ANALYTICAL FOUNDATION is committed to identify such personallities for their contributions across various field of Science and Technology and awarding them yearly. To participate for award, send us your details of research / testing / publication at <a href="mailto:info@analyticalfoundation.org">info@analyticalfoundation.org</a>
- 2. Improving quality of life by offering YOGA Training courses, Work shops / Seminars etc.
- 3. ANALYTICAL FOUNDATION aims to DETOXIFY human minds, souls and body by means of Yoga, Meditation, Ayurveda, Health Care, Awards, Media, Events, Camps etc.

# >>> Reach us @





HPLC Solutions MultipleLabs Analytical Bio-Med Analytical Distributors Analytical Foundation (Trust)

Company reserves rights to add/delete/modify the contents / technical specificationsof the catalogue without prior