

HS - 3200

Handheld Spectrometer



EPCC / PRODUCTS / APPLICATION / SOFTWARE / ACCESSORIES / CONSUMABLES / SERVICES

Analytical Technologies Limited

An ISO 9001 Certified Company


►► Specification

Spectrum		
Sensor	CMOS Linear Image Sensor	
Illuminance meter class	Directional response conforms to JIS C 1609-1:2006 for General Class AA. Directional response conforms to DIN 5032 Part 7 Class B.	
Wavelength Range	350 to 800 nm	
Wavelength Data Increment	1 nm	
Spectral Bandwidth	Approximately 9 nm (Half Bandwidth)	
Wavelength Reproducibility	± 1 nm ^{*1}	
Measurement Range	1. 70 ~ 150,000 lx 2. 0.5~1,000 W/m ² (Irradiance) 3. 1~3,000 μmol/(m ² *s) (PPFD)	
Illuminance Accuracy	Illuminant A @ 2,856 Kat 20,000 lx ^{*2}	± 5%
Illuminance Repeatability (2σ)		0.2%
Color Accuracy		x y : ± 0.0025
Color Repeatability (2σ)		x y : 0.0005
CCT Accuracy		± 2%
CRI Accuracy @ Ra		± 1.5%
Stray Light	-25 dB max. ^{*3}	
Integration Time Range	100 μs to 1,000 ms	
Digital Resolution	16 bits	

Feature	
Capture Function	One time / Continuous
Dark Mode	Auto Dark
Operation Mode	Standalone Mode / Bluetooth Mode ^{*4} / USB Mode (MSC Mode ^{*5} +PC connection)
Integration Mode	Auto/Manual
G sensor mode	Axial Displacement (x,y)
Measuring Modes	1. Basic Mode
	2. Spectrum Mode
	3. PPFD Mode
	4. PPFD Spectrum Mode (Including reference spectrum - Chlorophyll A, Chlorophyll B, Beta-carotene, Phytochrome A red, Phytochrome A far red)
	5. CIE 1931/1976 Chromaticity Mode
	6. Logging Mode
	7. Grid Mode
	8. Compare Mode
	9. Browser Mode
	10. Option Mode

Measuring Capabilities	1. Illuminance (LUX)/Foot Candle (fc)
	2. Correlated Color Temperature (CCT)
	3. CIE Chromaticity Coordinates (1) CIE 1931 x,y Coordinates (2) CIE 1976 u',v' Coordinates
	4. Δx , Δy , $\Delta u'$, $\Delta v'$
	5. Delta uv (Duv)
	6. Dominant Wavelength (λ_d)
	7. Excitation Purity
	8. Color Rendering Index (CRI, Ra)/R1 to R15
	9. Spectral Power Distribution (SPD)
	10. Peak Wavelength (λ_p)
	11. Peak Wavelength Value (λ_{pV})
	12. Intergration Time (I-Time)
	13. Irradiance (380nm~780nm)
	14. Photosynthetically Active Radiation (PAR) (1) PPFD(400nm~700nm) (2) PFD-R(600nm~700nm) (3) PFD-G(500nm~600nm) (4) PFD-B(400nm~500nm) (5) PFD(380nm~780nm) (6) PFD-UV(350nm~400nm) (7) PFD-FR(700nm~800nm) (8) PFD-B:G Ratio (400~500nm:500~600nm) (9) PFD-R:FR Ratio (600~700nm:700~800nm)
	15. Phytochrome Photostationary State(PSS)
	16. Temperature (°C)*7
	17. Relative Humidity (% RH)*7

System Configurations	
Display	4.3" 800X480 Capacitive Touch LCD
Waterproof level	IP66*6
Max. Files	≒ 68,000 Files @ 8GB SD Card (Excel + JPG)
Battery Operation Time	≦ 5 hours / Fully Charged
Power	Adapter; 3200 mAh (3.7V Rechargeable Li-ion Battery)
Data Output Interface	MicroSD Card (SD2.0,SDHC / up to 32G) / Type C / Bluetooth 3.0 and 4.0 compatible with iOS and Android
Data Format	Compatible Excel / JPG
Dimensions	190 x 81.7 x 29.5 mm (H x W x D)
Weight (with Battery)	280 g ± 10 g
Operating Temperature / Humidity	0 to 35 °C, relative humidity 70% or less without condensation
Storage Temperature / Humidity	-10 to 40 °C, relative humidity 70% or less without condensation
Display languages	English / Traditional Chinese / Simplified Chinese / Japanese / Spanish / German / French / Italian / Russian
PC Software	uSPECTRUM



- *1: Input source must be a stable light source.
- *2: Temperature $23 \pm 2^{\circ}\text{C}$ and relative humidity 50% or less.
- *3: Input the 550nm monochromatic light and measure the stray light ratio at $550\text{nm} \pm 40\text{nm}$.
- *4: It can be connected to mobile phones and tablets.
- *5: MSC- Mass Storage Class.
- *6: Only sensor, not the whole body
- *7: It has to be used with "PG200N Thermo-Hygro cable" to do the measurement.

HPLC Servicing, Validation, Trainings and Preventive Maintenance :

HPLC Servicing :HPLC Servicing : We have team of service engineers who can attend to any make of HPLC promptly @the most affordable cost.

Trainings :We also take up preventive Maintenance to reduce downtime of HPLC's Trainings.

AMC's/CMC :AMC's/CMC :We offer user training both in-House and at customer sites on HPLC principles, operations, trouble-shooting.

Validations :Validations :We have protocols for carrying out periodic Validations as per GLP/GMP/USFDA norms.

Instruments :Instruments :We offer instruments/Renting Services Modules like pumps,detector etc. on Rent.



About Analytical Technologies

Analytical Technologies is synonymous for offering technologies for doing analysis and is the Fastest Growing Global Brand having presence in at least 96 countries across the global. Analytical Technologies Limited is an ISO:9001 Certified Company engaged in Designing, Manufacturing, Marketing & providing Services for the Analytical, Chromatography, Spectroscopy, Bio Technology, Bio Medical, Clinical Diagnostics, Material Science & General Laboratory Instrumentation. Analytical Technologies, India has across the Country operations with at least 4 Regional Offices, 6 Branch Offices & Service Centers. Distributors & Channel partners worldwide.

Our Products & Technologies



UV/VIS
Spectro 2080+
Double Beam



Infra FTIR



Optima Gas
Chromatograph
3007



Optima Gas
Chromatograph
2979 Plus



Flash
Chromatograph



Atomic Absorption
Spectrophotometer



Liquid Partical
Counter



Optical Emission
Spectrophotometer



DSC/TGA



Semi Auto Bio
Chemistry Analyzer



HEMA 2062
Hematology
Analyzer



Micro Plate
Reader/Washer



URINOVA 2800
Urine Analyzer



Total Organic
Carbon 3800



Fully Automated
CLIA



NOVA-2100
Chemistry Analyzer



PCR/Gradient PCR/
RTPCR



TOC
Analyzer



Laser Particle
Size Analyzer



Ion Chromatograph



Water purification
system

Regulatory compliances



Corporate Social Responsibility

Analytical Foundation is a nonprofit organization (NGO) found for the purpose of:



Analytical
Foundation

1. Research & Innovation Scientist's awards/QC Professional Award : Quality life is possible by innovation only and the innovation is possible by research only, hence ANALYTICAL FOUNDATION is committed to identify such personalities for their contributions across various field of Science and Technology and awarding them yearly. To participate for award, send us your details of research / testing / publication at Info@analyticalfoundation.org

2. Improving quality of life by offering YOGA Training courses, Work shops/Seminars etc.

3. ANALYTICAL FOUNDATION aims to DETOXYFY human minds,souls and body by means of yoga, Meditation, Ayurveda, Health Care, Awards, Media, Events, Camps etc.



HPLC Solutions MultipleLabs Analytical Bio-Med Analytical Distributors Analytical Foundation (Trust)

Corporate & Regd. Office:
Analytical House, # E67 & E68,
Ravi Park, Vasna Road, Baroda,
Gujarat 390 015. INDIA

T: +91 265 2253620
+91 265 2252839
+91 265 2252370
F: +91 265 2254395

E: info@hplctechnologies.com
info@multiplelabs.com
info@analyticalgroup.net

W. www.analyticalgroup.net
www.hplctechnologies.com
www.multiplelabs.com

Sales & Support Offices:
across the country :
Distributors & Channel
partners World Wide